

Certificate of Analysis

No. CA 13082

Product: Eichrom's Anion Exchange Resin
Lot #: 060613A8M

Eichrom's Anion Exchange resin, Lot # 060613A8M has been manufactured according to Eichrom standard operating procedures. This lot has been tested in accordance with Eichrom quality assurance policy and procedures. This lot of product meets all Eichrom quality specifications.

ANALYSIS RESULTS

<u>Test Value</u>	<u>Specification</u>	<u>Result</u>
Appearance	White to light Yellow	Pass
Extractables	< 1 mg(residue) / g(resin)	0.01 mg/g
pH	≥ 4	Pass



10-Jun-13

Jill Bryant
Quality System Coordinator