

Certificate of Analysis

No. CA 15141


Product: Eichrom's Anion Exchange Resin
Lot #: 062415A8M

Eichrom's Anion Exchange resin, Lot # 062415A8M has been manufactured according to Eichrom standard operating procedures. This lot has been tested in accordance with Eichrom quality assurance policy and procedures. This lot of product meets all Eichrom quality specifications.

ANALYSIS RESULTS

<u>Test Value</u>	<u>Specification</u>	<u>Result</u>
Appearance	White to light Yellow	Pass
Extractables	< 1 mg (residue) / g (resin)	n.d.
pH	≥ 4	Pass
Wet Volume Capacity	≥ 1.2 meq/mL	1.5

Jill Bryant
Quality System Coordinator



July 21, 2015