

Certificate of Analysis


No. CA 17128

Product: Eichrom's Anion Exchange Resin
Date of Manufacture: 10-Aug-2017
Lot #: 072617A8M

Eichrom's Anion Exchange resin Lot 072617A8M has been manufactured according to Eichrom standard operating procedures. This lot has been tested in accordance with Eichrom quality assurance policy and procedures. This lot of product meets all Eichrom quality specifications.

ANALYSIS RESULTS

<u>Test Value</u>	<u>Specification</u>	<u>Result</u>
Appearance	White to light Yellow	Pass
Extractables	< 1 mg (residue) / g (resin)	0.07
pH	≥ 4	Pass
Wet Volume Capacity	≥ 1.2 meq/mL	1.2



Jill Bryant
Quality System Coordinator

August 11, 2017