

# Certificate of Analysis

No. CA 14035

Product: Eichrom's Anion Exchange Resin cartridges  
Lot #: A8MR14A

Eichrom's Anion Exchange resin cartridges, Lot # A8MR14A has been manufactured according to Eichrom standard operating procedures. The resin used to manufacture this lot has been tested in accordance with Eichrom quality assurance policy and procedures. This lot of product meets all Eichrom quality specifications.

## ANALYSIS RESULTS

<u>Test Value</u>	<u>Specification</u>	<u>Result</u>
Appearance	White to light Yellow	Pass
Extractables	< 1 mg(residue) /g (resin)	0.2
Wet Volume Capacity,	Minimum 1.2 meq/mL	1.4
Cl form		
pH	≥ 4	Pass

Jill Bryant  
Quality System Coordinator



March 6, 2014