

# Certificate of Analysis

No. CA 17036

Product: Eichrom's Anion Exchange Resin Cartridges  
Lot #: A8MR17C

Eichrom's Anion Exchange resin Cartridge Lot A8MR17C has been manufactured according to Eichrom standard operating procedures. The resin used to manufacture this lot, 120215A8M, has been tested in accordance with Eichrom quality assurance policy and procedures. This lot of product meets all Eichrom quality specifications.

## ANALYSIS RESULTS

<u>Test Value</u>	<u>Specification</u>	<u>Result</u>
Appearance	White to light Yellow	Pass
Extractables	< 1 mg (residue) / g (resin)	0.01
pH	≥ 4	Pass
Wet Volume Capacity	≥ 1.0 meq/mL	1.2

Jill Bryant  
Quality System Coordinator



March 9, 2017