

## Certificate of Analysis

No. CA09059

Product: Eichrom's Actinide Resin columns  
Lot #: AC09A

Eichrom's Actinide Resin Lot # AC09A has been manufactured according to Eichrom standard operating procedures. This resin lot has been tested in accordance with Eichrom quality assurance policy and procedures. This lot of product meets all Eichrom quality specifications.

### ANALYSIS RESULTS

<u>Test Value</u>	<u>Specification</u>	<u>Result</u>
Appearance	White beads	Pass
Distribution Ratio, 0.5 M HCl acid	$\geq 300$	393

*Jill Bryant*

May 11, 2009

Jill Bryant  
Quality System Coordinator