

## Certificate of Analysis


No. CA08171

Product: Eichrom's Actinide Resin columns  
Lot #: AC8A

Eichrom's Actinide Resin column Lot # AC8A has been manufactured according to Eichrom standard operating procedures. The resin used to manufacture this lot has been tested in accordance with Eichrom quality assurance policy and procedures. This lot of product meets all Eichrom quality specifications.

### ANALYSIS RESULTS

<u>Test Value</u>	<u>Specification</u>	<u>Result</u>
Appearance	White beads	Pass
Distribution Ratio, 0.5 M HCl acid	$\geq 300$	489



Jill Bryant  
Quality System Coordinator

August 21, 2008