

Certificate of Analysis

No. CA06057

Product: Eichrom's Actinide Resin

Lot #: ACA05196

Eichrom's Actinide Resin Lot # ACA05196 has been manufactured according to Eichrom standard operating procedures. This lot has been tested in accordance with Eichrom quality assurance policy and procedures. This lot of product meets all Eichrom quality specifications.

ANALYSIS RESULTS

| <u>Test Value</u> | <u>Specification</u> | <u>Result</u> |
|------------------------------------|----------------------|---------------|
| Appearance | White beads | Pass |
| Distribution Ratio, 0.5 M HCl acid | ≥ 300 | 426 |



Jill Bryant
Quality Control Chemist

Prepared: June 9, 2006