

Certificate of Analysis


No. CA10143

Product: Eichrom's Actinide Resin
Lot #: ACA071910

Eichrom's Actinide Resin Lot # ACA071910 has been manufactured according to Eichrom standard operating procedures. This resin lot has been tested in accordance with Eichrom quality assurance policy and procedures. This lot of product meets all Eichrom quality specifications.

ANALYSIS RESULTS

<u>Test Value</u>	<u>Specification</u>	<u>Result</u>
Appearance	White beads	Pass
Distribution Ratio, 0.5 M HCl acid	≥ 300	412



July 27, 2010

Jill Bryant
Quality System Coordinator