

Certificate of Analysis

No. CA07120

Product: Eichrom's Actinide Resin
Lot #: ACA08097

Eichrom's Actinide Resin Lot # ACA08097 has been manufactured according to Eichrom standard operating procedures. This lot has been tested in accordance with Eichrom quality assurance policy and procedures. This lot of product meets all Eichrom quality specifications.

ANALYSIS RESULTS

<u>Test Value</u>	<u>Specification</u>	<u>Result</u>
Appearance	White beads	Pass
Distribution Ratio, 0.5 M HCl acid	≥ 300	489



Jill Bryant
Quality Control Chemist

Prepared: August 23, 2007