



# Certificate of Analysis

No. CA 14240

Product: Eichrom's Cation Exchange Resin cartridges

Lot #: C8MR14F

Eichrom's Cation Exchange resin cartridges, Lot # C8MR14F has been manufactured according to Eichrom standard operating procedures. The resin used to manufacture this lot has been tested in accordance with Eichrom quality assurance policy and procedures. This lot of product meets all Eichrom quality specifications.

## ANALYSIS RESULTS

<u>Test Value</u>	<u>Specification</u>	<u>Result</u>
Appearance	Tan to orange-brown	Pass
Extractables	< 1 mg(residue) / g(resin)	0.25 mg/g
pH	≥ 4.0	pass

Jill Bryant  
Quality System Coordinator

December 11, 2014