

Product: Eichrom's DGA Resin-Normal
Lot #: DNS02277


Eichrom's DGA Resin-Normal Lot # DNS02277 has been manufactured according to Eichrom's standard operating procedures. This lot has been tested in accordance with Eichrom's quality assurance policy and procedures. This lot of product meets all Eichrom quality specifications.

ANALYSIS RESULTS

Test Value	Specification	Result
Appearance	White beads	Pass
Eu Breakthrough	< 2 % in Load	< 0.1 %
Eu Recovery	> 95 % in 3 BV	99.4 %

cpm/mL vs. Bed Volumes of Eluate

50-100 μm DGA Resin, normal, Preconditioned with 4.0 M HNO_3 , 22(1) $^\circ\text{C}$



Jill Bryant

Jill Bryant
Quality Control Chemist

Prepared: March 19, 2006