


Certificate of Analysis
No. CA11052

Product: Eichrom's DGA-Normal Resin
Lot #: DNS030111

Eichrom's DGA-Normal Resin Lot # DNS030111 has been manufactured according to Eichrom's standard operating procedures. This lot has been tested in accordance with Eichrom's quality assurance policy and procedures. This lot of product meets all Eichrom quality specifications.

ANALYSIS RESULTS

<u>Test Value</u>	<u>Specification</u>	<u>Result</u>
Appearance	White beads	Pass
% ²⁴¹ Am stripped in 0.1M HCl		
Least acceptable mean (4 rep)	97.0	99.2
Max. standard deviation	2.5	0.5
% ²⁴¹ Am stripped in 0.1M HNO ₃		
Maximum acceptable mean (4 rep)	1.0	0.15
Max. standard deviation	0.5	0.4
% ²³³ U stripped in 0.1M HNO ₃		
Least acceptable mean (4 rep)	97.0	98.7
Max. standard deviation	2.5	0.1
% ²³³ U stripped in 0.1M HCl		
Maximum acceptable mean (4 rep)	1.0	0.48
Max. standard deviation	0.5	0.1


Sarah Tejhima
Quality Supervisor

3-14-11
March 14, 2011