

**Certificate of Analysis**  
No. CA17063


Product: Eichrom's DGA-Normal Resin  
Lot #: DNS031617

Eichrom's DGA-Normal Resin Lot # DNS031617 has been manufactured according to Eichrom's standard operating procedures. This lot has been tested in accordance with Eichrom's quality assurance policy and procedures. This lot of product meets all Eichrom quality specifications.

**ANALYSIS RESULTS**

<u>Test Value</u>	<u>Specification</u>	<u>Result</u>
Appearance	White beads	Pass
% <sup>241</sup> Am stripped in 0.1M HCl		
Least acceptable mean (4 rep)	97.0	99.5
Max. standard deviation	2.5	0.5
% <sup>241</sup> Am stripped in 0.1M HNO <sub>3</sub>		
Maximum acceptable mean (4 rep)	1.0	0.1
Max. standard deviation	0.5	0.4
% <sup>233</sup> U stripped in 0.1M HNO <sub>3</sub>		
Least acceptable mean (4 rep)	97.0	99.1
Max. standard deviation	2.5	0.1
% <sup>233</sup> U stripped in 0.1M HCl		
Maximum acceptable mean (4 rep)	1.0	0.4
Max. standard deviation	0.5	0.0

Jill Bryant  
Quality System Coordinator



April 19, 2017