

Certificate of Analysis

No. CA08124


Product: Eichrom's DGA Resin-Normal
Lot #: DNS05078

Eichrom's DGA Resin-Normal Lot # DNS05078 has been manufactured according to Eichrom's standard operating procedures. This lot has been tested in accordance with Eichrom's quality assurance policy and procedures and meets all Eichrom quality specifications.

ANALYSIS RESULTS

Test Value	Specification	Result
Appearance	White beads	Pass
Eu Breakthrough	< 2 % in Load	0.1 %
Eu Recovery	> 95 %	97.1 %

DGA Resin Elution Curve



Jill Bryant

Jill Bryant
Quality Control Chemist

June 18, 2008