

Certificate of Analysis
No. CA09028


Product: Eichrom's DGA Resin-Normal cartridges
Lot #: DNSR09B

Eichrom's DGA Resin-Normal cartridge Lot # DNSR09B has been manufactured according to Eichrom's standard operating procedures. The resin used to manufacture this lot (DNS09098) has been tested in accordance with Eichrom's quality assurance policy and procedures and meets all Eichrom quality specifications.

ANALYSIS RESULTS

Test Value	Specification	Result
Appearance	White beads	Pass
Eu Breakthrough	< 2 % in Load	0.0 %
Eu Recovery	> 95 %	98.8 %

DGA Normal Elution Curve



Jill Bryant

March 23, 2009

Jill Bryant
Quality System Coordinator