

Certificate of Analysis
 No. CA09098


Product: Eichrom's DGA Resin-Normal cartridge
 Lot #: DNSR09C

Eichrom's DGA Resin-Normal cartridge Lot # DNSR09C has been manufactured according to Eichrom's standard operating procedures. This lot has been tested in accordance with Eichrom's quality assurance policy and procedures and meets all Eichrom quality specifications.

ANALYSIS RESULTS

<u>Test Value</u>	<u>Specification</u>	<u>Result</u>
Appearance	White beads	Pass
Eu Breakthrough	< 2 % in Load	0.0 %
Eu Recovery	> 95 %	99.4 %

DGA Normal Elution Curve



Jill Bryant
 Quality System Coordinator

August 3, 2009