

**Certificate of Analysis**  
No. CA10180


Product: Eichrom's DGA-Normal Resin 1mL cartridges  
Lot #: DNSR10D

Eichrom's DGA-Normal Resin 1 mL cartridges Lot # DNSR10D has been manufactured according to Eichrom's standard operating procedures. The resin used to manufacture this lot has been tested in accordance with Eichrom's quality assurance policy and procedures. This lot of product meets all Eichrom quality specifications.

**ANALYSIS RESULTS**

<u>Test Value</u>	<u>Specification</u>	<u>Result</u>
Appearance	White beads	Pass
% <sup>241</sup> Am stripped in 0.1M HCl		
Least acceptable mean (4 rep)	97.0	98.1
Max. standard deviation	2.5	0.3
% <sup>241</sup> Am stripped in 0.1M HNO <sub>3</sub>		
Maximum acceptable mean (4 rep)	1.0	0.5
Max. standard deviation	0.5	0.3
% <sup>233</sup> U stripped in 0.1M HNO <sub>3</sub>		
Least acceptable mean (4 rep)	97.0	98.6
Max. standard deviation	2.5	0.1
% <sup>233</sup> U stripped in 0.1M HCl		
Maximum acceptable mean (4 rep)	1.0	0.5
Max. standard deviation	0.5	0.0

Jill Bryant  
Quality System Coordinator



September 17, 2010