


## Certificate of Analysis No. CA07128

Product: Eichrom's DGA Resin-Normal cartridges  
Lot #: DNSR7B

Eichrom's DGA Resin-Normal cartridge Lot # DNSR7B has been manufactured according to Eichrom's standard operating procedures. The resin used to manufacture this lot has been tested in accordance with Eichrom's quality assurance policy and procedures. This lot of product meets all Eichrom quality specifications.

### ANALYSIS RESULTS

| Test Value      | Specification  | Result  |
|-----------------|----------------|---------|
| Appearance      | White beads    | Pass    |
| Eu Breakthrough | < 2 % in Load  | < 0.1 % |
| Eu Recovery     | > 95 % in 3 BV | 99.4 %  |



*Jill Bryant*

Jill Bryant  
Quality Control Chemist

Prepared: September 6, 2007