

**Certificate of Analysis**  
No. CA08138


Product: Eichrom's DGA Resin-Normal cartridges  
Lot #: DNSR8E

Eichrom's DGA Resin-Normal cartridge Lot # DNSR8E has been manufactured according to Eichrom's standard operating procedures. The resin used to manufacture this lot has been tested in accordance with Eichrom's quality assurance policy and procedures and meets all Eichrom quality specifications.

**ANALYSIS RESULTS**

<u>Test Value</u>	<u>Specification</u>	<u>Result</u>
Appearance	White beads	Pass
Eu Breakthrough	< 2 % in Load	0.1 %
Eu Recovery	> 95 %	97.1 %

**DGA Resin Elution Curve**



*Jill Bryant*

July 9, 2008

Jill Bryant  
Quality System Coordinator