

Certificate of Analysis

No. CA08200


Product: Eichrom's Tritium Columns

Lot #: H38A

Eichrom's Tritium Columns, lot# H38A has been manufactured according to Eichrom standard operating procedures. This lot has been tested in accordance with Eichrom quality assurance policy and procedures. This lot of product meets all Eichrom quality specifications.

ANALYSIS RESULTS

<u>Test Value</u>	<u>Specification</u>	<u>Result</u>
Radiometric yield of Tritium	90-110%	105 %



October 15, 2008

Jill Bryant
Quality System Coordinator