

Certificate of Analysis

No. CA17160


Product: Eichrom's LN-2 Resin Cartridges
 Lot #: L2SR17A
 Date of Manufacture: 26-SEP-2017

Eichrom's LN-2 Resin cartridge Lot # L2SR17A has been manufactured according to Eichrom standard operating procedures. The resin used to manufacture this lot, L2S091817, has been tested in accordance with Eichrom quality assurance policy and procedures. This lot of product meets all Eichrom quality specifications.

ANALYSIS RESULTS

<u>Test Value</u>	<u>Specification</u>	<u>Result</u>
Appearance	White Beads	Pass
Lu Peak Maximum	13.5-15.5 BV	14.5 BV
Yb Peak Maximum	8.0-9.0 BV	8.5 BV
Seperation Factor	1.6-1.8	1.7

Jill Bryant
 Quality System Coordinator



September 27, 2017