

Certificate of Analysis


No. CA10077

Product: Eichrom's LN Resin
Lot #: LNA042710

Eichrom's LN Resin Lot # LNA042710 has been manufactured according to Eichrom standard operating procedures. This lot has been tested in accordance with Eichrom quality assurance policy and procedures. This lot of product meets all Eichrom quality specifications.

ANALYSIS RESULTS

<u>Test Value</u>	<u>Specification</u>	<u>Result</u>
Appearance	White beads	Pass
Distribution Ratio, 0.1 N nitric acid	>700	1479



Jill Bryant
Quality System Coordinator

April 30, 2010