

Certificate of Analysis


No. CA10084

Product: Eichrom's LN Resin cartridges
Lot #: LNSR10A

Eichrom's LN Resin cartridge Lot # LNSR10A has been manufactured according to Eichrom standard operating procedures. The resin used to manufacture this lot has been tested in accordance with Eichrom quality assurance policy and procedures. This lot of product meets all Eichrom quality specifications.

ANALYSIS RESULTS

<u>Test Value</u>	<u>Specification</u>	<u>Result</u>
Appearance	White beads	Pass
Distribution Ratio, 0.1 N nitric acid	>700	2922



Jill Bryant
Quality System Coordinator

May 7, 2010