



# Certificate of Analysis

No. CA15002

Product: Eichrom's MnO<sub>2</sub> Resin

Lot #: MNA010915

Eichrom's MnO<sub>2</sub> Resin Lot # MNA010915 has been manufactured according to Eichrom's standard operating procedures. This lot has been tested in accordance with Eichrom's quality assurance policy and procedures. This lot of product meets all Eichrom quality specifications.

## ANALYSIS RESULTS

<u>Test Value</u>	<u>Specification</u>	<u>Result</u>
Ba-133 recovery (4 reps)	$\geq 95\%$	103.5 %
Ba-133 in Load solution (4 reps)	$< 1.0\%$	0.00 %

Sarah Teichma  
Quality Manager

16 Jan 15

January 16, 2015