

## Certificate of Analysis

No. CA12094

Product: Eichrom's MnO<sub>2</sub> Resin

Lot #: MNA051712

Eichrom's MnO<sub>2</sub> Resin Lot # MNA051712 has been manufactured according to Eichrom's standard operating procedures. This lot has been tested in accordance with Eichrom's quality assurance policy and procedures. This lot of product meets all Eichrom quality specifications.

### ANALYSIS RESULTS

<u>Test Value</u>	<u>Specification</u>	<u>Result</u>
Ba-133 recovery (4 reps)	≥ 95%	99.8 %
Ba-133 in Load solution (4 reps)	<1.0%	0.00 %

  
Sarah Tejchma  
QC Supervisor

5-24-12  
May 24, 2012