

Certificate of Analysis

No. CA11144


Product: Eichrom's MnO₂ Resin

Lot #: MNA052511

Eichrom's MnO₂ Resin Lot # MNA052511 has been manufactured according to Eichrom's standard operating procedures. This lot has been tested in accordance with Eichrom's quality assurance policy and procedures. This lot of product meets all Eichrom quality specifications.

ANALYSIS RESULTS

<u>Test Value</u>	<u>Specification</u>	<u>Result</u>
Ba-133 recovery (4 reps)	≥ 95%	97.1 %
Ba-133 in Load solution (4 reps)	<1.0%	0.85 %



June 23, 2011

Jill Bryant
Quality System Coordinator