

Certificate of Analysis

No. CA10102


Product: Eichrom's MnO_2 Resin

Lot #: MNA052610

Eichrom's MnO_2 Resin Lot # MNA052610 has been manufactured according to Eichrom's standard operating procedures. This lot has been tested in accordance with Eichrom's quality assurance policy and procedures. This lot of product meets all Eichrom quality specifications.

ANALYSIS RESULTS

Test Value	Specification	Result
Ba-133 recovery	$\geq 95\%$	99.6 %



May 28, 2010

Joel Williamson
Quality Manager