

Certificate of Analysis

No. CA12180

Product: Eichrom's MnO₂ Resin

Lot #: MNA090712

Eichrom's MnO₂ Resin Lot # MNA090712 has been manufactured according to Eichrom's standard operating procedures. This lot has been tested in accordance with Eichrom's quality assurance policy and procedures. This lot of product meets all Eichrom quality specifications.

ANALYSIS RESULTS

<u>Test Value</u>	<u>Specification</u>	<u>Result</u>
Ba-133 recovery (4 reps)	≥ 95%	98.7 %
Ba-133 in Load solution (4 reps)	<1.0%	0.72 %


Sarah Tejchma
QC Supervisor

9-24-12
September 24, 2012