

Certificate of Analysis


No. CA09149

Product: Eichrom's MnO_2 Resin
Lot #: MNA091809

Eichrom's MnO_2 Resin Lot # MNA091809 has been manufactured according to Eichrom's standard operating procedures. This lot has been tested in accordance with Eichrom's quality assurance policy and procedures. This lot of product meets all Eichrom quality specifications.

ANALYSIS RESULTS

Test Value	Specification	Result
Ba-133 recovery	$\geq 95\%$	97.0 %



Joel Williamson
Quality Manager

October 26, 2009