

Certificate of Analysis

No. CA10210

Product: Eichrom's MnO₂ Resin

Lot #: MNA100410

Eichrom's MnO₂ Resin Lot # MNA100410 has been manufactured according to Eichrom's standard operating procedures. This lot has been tested in accordance with Eichrom's quality assurance policy and procedures. This lot of product meets all Eichrom quality specifications.

ANALYSIS RESULTS

<u>Test Value</u>	<u>Specification</u>	<u>Result</u>
Ba-133 recovery (4 reps)	≥ 95%	98.7 %
Ba-133 in Load solution (4 reps)	<1.0%	0.0%



October 7, 2010

Joel Williamson
Quality Manager