

Certificate of Analysis

No. CA10263


Product: Eichrom's MnO₂ Resin

Lot #: MNA122110

Eichrom's MnO₂ Resin Lot # MNA122110 has been manufactured according to Eichrom's standard operating procedures. This lot has been tested in accordance with Eichrom's quality assurance policy and procedures. This lot of product meets all Eichrom quality specifications.

ANALYSIS RESULTS

<u>Test Value</u>	<u>Specification</u>	<u>Result</u>
Ba-133 recovery (4 reps)	≥ 95%	105.4 %
Ba-133 in Load solution (4 reps)	<1.0%	0.0%



Jill Bryant
Quality System Coordinator

December 30, 2010