

Certificate of Analysis

No. CA11189


Product: Eichrom's MnO_2 Resin cartridges

Lot #: MNAR11C

Eichrom's MnO_2 Resin cartridge Lot # MNAR11C has been manufactured according to Eichrom's standard operating procedures. The resin used to manufacture this lot has been tested in accordance with Eichrom's quality assurance policy and procedures. This lot of product meets all Eichrom quality specifications.

ANALYSIS RESULTS

<u>Test Value</u>	<u>Specification</u>	<u>Result</u>
Ba-133 recovery (4 reps)	$\geq 95\%$	97.1 %
Ba-133 in Load solution (4 reps)	$< 1.0\%$	0.85 %



August 29, 2011

Jill Bryant
Quality System Coordinator