

Certificate of Analysis


No. CA08160

Product: Eichrom's MnO_2 Resin cartridges
Lot #: MNAR8A

Eichrom's MnO_2 Resin cartridge Lot # MNAR8A has been manufactured according to Eichrom's standard operating procedures. The resin used to manufacture this lot has been tested in accordance with Eichrom's quality assurance policy and procedures. This lot of product meets all Eichrom quality specifications.

ANALYSIS RESULTS

Test Value	Specification	Result
Ba-133 recovery	$\geq 95\%$	95%



Jill Bryant

August 19, 2008

Jill Bryant
Quality System Coordinator