

Certificate of Analysis


No. CA 17016

Product: Eichrom's Nickel Resin
Lot #: NIA020817

Eichrom's Nickel Resin Lot # NIA020817 has been manufactured according to Eichrom's standard operating procedures. This lot has been tested in accordance with Eichrom's quality assurance policy and procedures. This lot of product meets all Eichrom quality specifications.

ANALYSIS RESULTS

<u>Test Value</u>	<u>Specification</u>	<u>Result</u>
Appearance	White beads	Pass
Ave. % Ni Recovery	80-105%	98.1 %
Ave. % Ni in Load/Rinse	<1%	0.3 %
Ave. % Ni in Strip	>99%	99.7 %


Sarah McAlister
Quality Manager


February 13, 2017