

Certificate of Analysis

No. CA 15042


Product: Eichrom's Nickel Resin

Lot #: NIA022515

Eichrom's Nickel Resin Lot # NIA022515 has been manufactured according to Eichrom's standard operating procedures. This lot has been tested in accordance with Eichrom's quality assurance policy and procedures. This lot of product meets all Eichrom quality specifications.

ANALYSIS RESULTS

<u>Test Value</u>	<u>Specification</u>	<u>Result</u>
Appearance	White beads	Pass
Ave. % Ni Recovery	80-105%	94.4 %
Ave. % Ni in Load/Rinse	<1%	0.1 %
Ave. % Ni in Strip	>99%	99.9 %



Jill Bryant
Quality System Coordinator

March 13, 2015