

Certificate of Analysis


No. CA10038

Product: Eichrom's Nickel Resin
Lot #: NIA030210

Eichrom's Nickel Resin Lot # NIA030210 has been manufactured according to Eichrom's standard operating procedures. This lot has been tested in accordance with Eichrom's quality assurance policy and procedures. This lot of product meets all Eichrom quality specifications.

ANALYSIS RESULTS

| <u>Test Value</u> | <u>Specification</u> | <u>Result</u> |
|-------------------------|----------------------|---------------|
| Appearance | White beads | Pass |
| Ave. % Ni Recovery | 80-105% | 102.7% |
| Ave. % Ni in Load/Rinse | <1% | 0 % |
| Ave. % Ni in Strip | >99% | 100 % |



Joel Williamson
Quality Manager

March 5, 2010