

Certificate of Analysis

No. CA 14072


Product: Eichrom's Nickel Resin

Lot #: NIA040914

Eichrom's Nickel Resin Lot # NIA040914 has been manufactured according to Eichrom's standard operating procedures. This lot has been tested in accordance with Eichrom's quality assurance policy and procedures. This lot of product meets all Eichrom quality specifications.

ANALYSIS RESULTS

<u>Test Value</u>	<u>Specification</u>	<u>Result</u>
Appearance	White beads	Pass
Ave. % Ni Recovery	80-105%	96.3 %
Ave. % Ni in Load/Rinse	<1%	0.3 %
Ave. % Ni in Strip	>99%	99.7 %



Sarah Teichma
Quality Manager

30 Apr 14

April 30, 2014