

## Certificate of Analysis


No. CA09070

Product: Eichrom's Nickel Resin  
Lot #: NIA051809

Eichrom's Nickel Resin Lot # NIA051809 has been manufactured according to Eichrom's standard operating procedures. This lot has been tested in accordance with Eichrom's quality assurance policy and procedures. This lot of product meets all Eichrom quality specifications.

### ANALYSIS RESULTS

<u>Test Value</u>	<u>Specification</u>	<u>Result</u>
Appearance	White beads	Pass
Ave. % Ni Recovery	80-105%	103 %
Ave. % Ni in Load/Rinse	<1%	0.1%
Ave. % Ni in Strip	>99%	99.9%



June 1, 2009

Jill Bryant  
Quality System Coordinator