

## Certificate of Analysis

No. CA08119

Product: Eichrom's Nickel Resin

Lot #: NIA05278

Eichrom's Nickel Resin Lot # NIA05278 has been manufactured according to Eichrom's standard operating procedures. This lot has been tested in accordance with Eichrom's quality assurance policy and procedures. This lot of product meets all Eichrom quality specifications.

### ANALYSIS RESULTS

<u>Test Value</u>	<u>Specification</u>	<u>Result</u>
Appearance	White beads	Pass
Ave. % Ni Recovery	80-105%	98.4 %
Ave. % Ni in Load/Rinse	<1%	0%
Ave. % Ni in Strip	>99%	100%



June 13, 2008

Jill Bryant  
Quality Control Chemist