

Certificate of Analysis

No. CA08126

Product: Eichrom's Nickel Resin

Lot #: NIA06038

Eichrom's Nickel Resin Lot # NIA06038 has been manufactured according to Eichrom's standard operating procedures. This lot has been tested in accordance with Eichrom's quality assurance policy and procedures. This lot of product meets all Eichrom quality specifications.

ANALYSIS RESULTS

| <u>Test Value</u> | <u>Specification</u> | <u>Result</u> |
|-------------------------|----------------------|---------------|
| Appearance | White beads | Pass |
| Ave. % Ni Recovery | 80-105% | 99.3 % |
| Ave. % Ni in Load/Rinse | <1% | 0% |
| Ave. % Ni in Strip | >99% | 100% |



June 18, 2008

Jill Bryant
Quality Control Chemist