



Certificate of Analysis


No. CA 17121

Product: Eichrom's Nickel Resin
Date of Manufacture: 25-Jul-17
Lot #: NIA072517

Eichrom's Nickel Resin Lot # NIA072517 has been manufactured according to Eichrom's standard operating procedures. This lot has been tested in accordance with Eichrom's quality assurance policy and procedures. This lot of product meets all Eichrom quality specifications.

ANALYSIS RESULTS

<u>Test Value</u>	<u>Specification</u>	<u>Result</u>
Appearance	White beads	Pass
Ave. % Ni Recovery	80-105%	99.5 %
Ave. % Ni in Load/Rinse	<1%	0.1 %
Ave. % Ni in Strip	>99%	99.9 %



Jill Bryant
Quality System Coordinator

July 27, 2017

