

Certificate of Analysis


No. CA09099

Product: Eichrom's Nickel Resin
Lot #: NIA080409

Eichrom's Nickel Resin Lot # NIA080409 has been manufactured according to Eichrom's standard operating procedures. This lot has been tested in accordance with Eichrom's quality assurance policy and procedures. This lot of product meets all Eichrom quality specifications.

ANALYSIS RESULTS

| <u>Test Value</u> | <u>Specification</u> | <u>Result</u> |
|-------------------------|----------------------|---------------|
| Appearance | White beads | Pass |
| Ave. % Ni Recovery | 80-105% | 101 % |
| Ave. % Ni in Load/Rinse | <1% | 0 % |
| Ave. % Ni in Strip | >99% | 100 % |



Jill Bryant
Quality System Coordinator

August 12, 2009