

Certificate of Analysis

No. CA11199


Product: Eichrom's Nickel Resin
Lot #: NIA082411

Eichrom's Nickel Resin Lot # NIA082411 has been manufactured according to Eichrom's standard operating procedures. This lot has been tested in accordance with Eichrom's quality assurance policy and procedures. This lot of product meets all Eichrom quality specifications.

ANALYSIS RESULTS

<u>Test Value</u>	<u>Specification</u>	<u>Result</u>
Appearance	White beads	Pass
Ave. % Ni Recovery	80-105%	91.7 %
Ave. % Ni in Load/Rinse	<1%	0 %
Ave. % Ni in Strip	>99%	100 %

Jill Bryant
Quality System Coordinator



September 13, 2011