

Certificate of Analysis

No. CA 15175


Product: Eichrom's Nickel Resin
Lot #: NIA090815

Eichrom's Nickel Resin Lot # NIA090815 has been manufactured according to Eichrom's standard operating procedures. This lot has been tested in accordance with Eichrom's quality assurance policy and procedures. This lot of product meets all Eichrom quality specifications.

ANALYSIS RESULTS

<u>Test Value</u>	<u>Specification</u>	<u>Result</u>
Appearance	White beads	Pass
Ave. % Ni Recovery	80-105%	97.9 %
Ave. % Ni in Load/Rinse	<1%	0.1 %
Ave. % Ni in Strip	>99%	99.9 %

Jill Bryant
Quality System Coordinator



September 22, 2015