

Certificate of Analysis


No. CA 15197

Product: Eichrom's Nickel Resin
Lot #: NIA100715

Eichrom's Nickel Resin Lot # NIA100715 has been manufactured according to Eichrom's standard operating procedures. This lot has been tested in accordance with Eichrom's quality assurance policy and procedures. This lot of product meets all Eichrom quality specifications.

ANALYSIS RESULTS

<u>Test Value</u>	<u>Specification</u>	<u>Result</u>
Appearance	White beads	Pass
Ave. % Ni Recovery	80-105%	96.1 %
Ave. % Ni in Load/Rinse	<1%	0.1 %
Ave. % Ni in Strip	>99%	99.9 %



Sarah McAlister
Quality Manager

12 Oct 15
October 12, 2015