

## Certificate of Analysis

No. CA12199

Product: Eichrom's Nickel Resin


Lot #: NIA102212

Eichrom's Nickel Resin Lot # NIA102212 has been manufactured according to Eichrom's standard operating procedures. This lot has been tested in accordance with Eichrom's quality assurance policy and procedures. This lot of product meets all Eichrom quality specifications.

### ANALYSIS RESULTS

<u>Test Value</u>	<u>Specification</u>	<u>Result</u>
Appearance	White beads	Pass
Ave. % Ni Recovery	80-105%	94.8%
Ave. % Ni in Load/Rinse	<1%	0.3 %
Ave. % Ni in Strip	>99%	99.7 %

Jill Bryant  
Quality System Coordinator



October 25, 2012