

Certificate of Analysis


No. CA11230

Product: Eichrom's Nickel Resin
Lot #: NIA102511

Eichrom's Nickel Resin Lot # NIA102511 has been manufactured according to Eichrom's standard operating procedures. This lot has been tested in accordance with Eichrom's quality assurance policy and procedures. This lot of product meets all Eichrom quality specifications.

ANALYSIS RESULTS

<u>Test Value</u>	<u>Specification</u>	<u>Result</u>
Appearance	White beads	Pass
Ave. % Ni Recovery	80-105%	90.8 %
Ave. % Ni in Load/Rinse	<1%	0 %
Ave. % Ni in Strip	>99%	100 %



Sarah Fejchma
QC Supervisor

10-28-11
October 28, 2011