

Certificate of Analysis


No. CA10236

Product: Eichrom's Nickel Resin
Lot #: NIA110510

Eichrom's Nickel Resin Lot # NIA110510 has been manufactured according to Eichrom's standard operating procedures. This lot has been tested in accordance with Eichrom's quality assurance policy and procedures. This lot of product meets all Eichrom quality specifications.

ANALYSIS RESULTS

<u>Test Value</u>	<u>Specification</u>	<u>Result</u>
Appearance	White beads	Pass
Ave. % Ni Recovery	80-105%	104 %
Ave. % Ni in Load/Rinse	<1%	0 %
Ave. % Ni in Strip	>99%	100 %



Jill Bryant

Quality System Coordinator

November 11, 2010