

Certificate of Analysis


No. CA11277

Product: Eichrom's Nickel Resin
Lot #: NIA121211

Eichrom's Nickel Resin Lot # NIA121211 has been manufactured according to Eichrom's standard operating procedures. This lot has been tested in accordance with Eichrom's quality assurance policy and procedures. This lot of product meets all Eichrom quality specifications.

ANALYSIS RESULTS

<u>Test Value</u>	<u>Specification</u>	<u>Result</u>
Appearance	White beads	Pass
Ave. % Ni Recovery	80-105%	102 %
Ave. % Ni in Load/Rinse	<1%	0 %
Ave. % Ni in Strip	>99%	100 %


Sarah Teichma
QC Supervisor

12-22-11
December 22, 2011